

# Computer-Aided Detection False Positives

Subjects: Gastroenterology & Hepatology

Submitted by:  Yu-Hsi

Hsieh

## Definition

Defining FPs based on the duration of time is an objective way of classifying FPs. However, the threshold required for reporting FPs is unsettled. One report suggested that only FPs > 2 s be reported, and another only reported FPs > 1 s, while the majority of FPs (i.e., more than 90%) lasted <0.5 s. It is unknown whether ignoring the transient FPs (i.e., those lasting for <1 or 2 s) would increase the risk of missing a real polyp.

---

## 1. Introduction

Missed lesions account for 57.8% of interval colorectal cancers (i.e., cancers that occur within 3–5 years after a negative colonoscopy) [1]. To reduce incidences of missed lesions and interval cancers, measures were proposed to improve the quality of colonoscopies. One of the most important quality metrics is the adenoma detection rate (ADR), defined as the proportion of patients with at least one adenoma [2].

Artificial intelligence (AI) is being used in the computer-aided detection (CADE) and diagnosis (CADx) of polyps [3]. Randomized controlled trials (RCTs) showed CADE-assisted colonoscopy significantly increased the ADR [4][5][6][7][8]. A meta-analysis confirmed that the ADR was significantly higher in the CADE group than in the conventional group (36.6% vs. 25.2%; RR, 1.44; 95% confidence interval, 1.27–1.62;  $p < 0.01$ ;  $I^2 = 42\%$ ) [9].

An accompanying limitation of the CADE is false positives (FPs), which occur when the algorithm identifies a “polyp” that the endoscopist would disagree with. FPs were ranked 3rd in importance among 59 future research questions related to CADE [10]. We assessed CADE-overlaid video analyses, RCTs using real-time CADE to enhance polyp detection during colonoscopies, and studies that used FPs as the primary outcome. We test the hypothesis that the systematic review of the literature on FPs will yield insight into methods of managing and limiting the adverse effects of this drawback of CADE.

## 2. Adverse Effects of FPs

The time expended to differentiate an FP from a true lesion can potentially increase the withdrawal time. Although most RCTs on the real-time application of CADE found a longer withdrawal time in the CADE group compared to the control group [4][6][7][8], the withdrawal time without biopsy was not significantly different. In a post hoc analysis of a small fraction (40/342 or <11.7%) of the original CADE groups in the RCT studies, Hassan et al. found that 94% of FPs were discarded by the endoscopist immediately without further exploration, and the time wasted on the remaining FPs only contributed to about 1% of the withdrawal time. In a real-life situation, where the bowel preparation is usually less than optimal and endoscopists are less experienced, the impacts of bowel preparation on FPs and withdrawal time require more objective studies.

The presence of FPs might lead to unnecessary biopsies of non-neoplastic tissues. (with another 1 unreported [6] and 1 showing no difference [5]) listed in **Table 1** showed a significant increase in the biopsy of non-neoplastic polyps in the CADE group, which was typically double the number reported for the control group. The removal of hyperplastic polyps—other than the diminutive ones at the distal rectosigmoid colon—is justified, as these polyps contribute to the serrated pathway of colorectal carcinogenesis [11]. If these biopsies were, in fact, unwarranted, then there exists an avoidable non-indicated use of medical resources.

**Table 1.** Recent RCTs comparing real-time CADE with control on adenoma detection during colonoscopy.

Study	Location of Study	Control vs. CAde (n)	Overall ADR	Non-Neoplastic Polyps Detected, n (%)	CAde Used During Insertion	Number of Screens Used	Withdrawal Time, Mean, Minutes	Withdrawal Time, Exclude Biopsy, Mean, Minutes
Wang et al. [7]	China	536 vs. 522	20.3% vs. 29.1% *	94 (34.9) vs. 217 (43.6) * (hyperplastic plus inflammatory)	No	2	6.39 vs. 6.89 *	6.07 vs. 6.18
Wang et al. [8]	China	478 vs. 484	28% vs. 34% *	113 (37) vs. 200 (40) * (hyperplastic plus inflammatory)	No	1	6.99 vs. 7.46 *	6.37 vs. 6.48
Repici et al. [5]	Italy	344 vs. 341	40.4% vs. 54.8% *	57 (16.6) vs. 68 (19.9) (Normal, hyperplastic, inflammatory and others)	Yes	1	NA	7.0 vs. 7.3
Su et al. [6]	China	308 vs. 315	16.5% vs. 28.9% *	NA	No	2	5.68 vs. 7.03 *	6.74 vs. 6.82 *
Liu et al. [4]	China	518 vs. 508	23.9% vs. 39.1% *	92 (37.1) vs. 203 (41.8) * (proliferative and inflammatory)	No	2	NA	6.32 vs. 6.37
Liu et al. [12]	China	397 vs. 393	20.9% vs. 29.0% *	87 (42.7) vs. 222 (52.7) * (hyperplastic and inflammatory)	No	1	6.94 vs. 7.29 *	6.62 vs. 6.71

ADR, adenoma detection rate; CAde, computer-aided detection; FP, false positive; FPR, false positive rate; NA, not analyzed; RCT, randomized controlled trial; \*, statistically significant; NA, not analyzed.

The application of the CADx to characterize the polyps following their detection with the CAde might help reduce the number of unnecessary polypectomies of non-neoplastic polyps. Preliminary results showed promise for simultaneously classifying polyps with endocytoscopic images [13], or even with white light images [14] after using the CAde to detect the polyps in white light.

The recurrent appearance of FPs on the screen may lead to increased fatigue and decreased vigilance on the part of the endoscopist [15]. Inundating the endoscopist with such a large amount of prompts on the screen, even if only very transient attention is demanded for each prompt, engenders the risk of the fatigue of the endoscopist. However, a study showed that a real-time CAde system, integrated on one primary endoscopy monitor instead of the two monitors used in most RCTs (**Table 1**), improved the ADR without an increase in the subjective fatigue level reported by the endoscopists during the colonoscopy [12]. The unblinded report, developed by proponents of the CAde algorithm under study, raised questions regarding the objectivity of the results.

False positives cause distractions and the need for refocusing, potentially resulting in adverse effects during the search for real polyps. To illustrate how difficult it is to refocus after distraction, a study on mobile phone use while driving showed that the risk of a rear-end accident occurring increased by 2.34–3.56 times, despite increasing their time headway by 0.41–0.59 s to offset the distraction of texting while driving [16].

Too many FPs may hamper the enthusiasm of the endoscopist to apply the CAde in clinical practice. One recent survey on the views of gastroenterologists regarding the potential use of artificial intelligence found that 33.9% of respondents worried about high numbers of FPs [17]. Reports that emphasize the lack

of importance of FPs based on subjective assessment need to be re-evaluated by studies with more objective and unbiased designs.

### 3. How to Address the Occurrence of FPs

There is considerable variability in FPRs in the literature (**Table 2**). This variability suggests that there are diverse definitions of FPs and various conditions that affect the occurrence of FPs inside the bowel lumen, which indicates that there is an opportunity to minimize FPs through standardizing the definitions of FPs and optimizing the condition of the bowel lumen.

**Table 2.** Recent studies using CADe-overlaid videos for real-time detection of polyps.

Study	Primary Outcome	Videos Reviewed (n)	Polyps Detected	Sensitivity	Specificity
Misawa et al. [18]	Accuracy of CADe	155 positive videos and 391 negative videos. Most of the polyps were flat.	NA	Per-frame: 90%	Per-frame: 63.3%
Urban et al. [19]	Polyp detection by CADe	9 randomly selected colonoscopy videos	Performing endoscopist: 28 Three expert reviewers without CADe: 36 One expert reviewer with CADe: 45	Per-polyp: 94%	Per-frame: 93%
Becq et al. [20]	Polyp detection by CADe	50 colonoscopies from consecutive patients with various bowel preparations.	Performing endoscopist: 55 CADe: 401 possible polyps (100 definite polyps, 63 possible polyps, and 238 false positives)	Per-polyp: 98.8%	NA
Guo et al. [21]	Accuracy of CADe	50 videos with small polyps and 50 videos without polyps.	NA	When confidence level $\geq 10\%$ , per-frame: 66.9% When confidence level $\geq 30\%$ , per-frame: 56.8%	When confidence level $\geq 10\%$ , per-frame: 92% When confidence level $\geq 30\%$ , per-frame: 98%
Wang et al. [22]	Accuracy of CADe	138 videos with polyps and 54 videos without polyps	NA	Per-frame: 91.6%	Per-frame: 95.4%
Misawa et al. [23]	Accuracy of CADe in a large, publicly accessible database.	100 videos	NA	Per-frame: 90.5% Per-polyp: 98.0%	Per frame: 93.7%
Hassan et al. [24]	Accuracy of CADe	138 polyp-positive short videos	NA	Per-frame: 99.7%	NA
Lee et al. [25]	Accuracy of CADe	15 unaltered videos	Performing endoscopist: 38 CADe: 45	Per-frame: 89.3%	NA
Podlasek et al. [26]	Accuracy of CADe	42 colonoscopy videos	Reviewer: 84 CADe: 79	Per-polyp: 94.1%	NA

CADe: computer-aided detection; NA, not analyzed.

An example of a simple method that could be used to reduce FPs is re-training the CADe algorithms with scenarios that currently lead to FPs. Another approach could be the adoption of recurrent neural networks, which have memory and can process temporal sequences of frames in a way that is similar to

the learning process of human brains [10]. (You Only Look Once, Version 3), a state-of-the-art, real-time object detection algorithm, better specificity was achieved (increasing from 90.9% to 93.7%) To filter out most short flashes, Podlasek et al. suggested setting a threshold of persistent time for FPs to show up; however, this method might introduce a minor detection lag, depending on the desired sensitivity [26].

Optimal bowel preparation is the prerequisite for a high-quality CADe-assisted colonoscopy and is associated with fewer FPs [27]. As the major source of CADe FP alerts is the wrinkled walls, they can be reduced by ensuring adequate luminal insufflation. The use of an anti-spasmodic agent, such as Hyoscine-n-butylbromide, might be helpful in reducing the contraction of the colon wall [28]. Adding simethicone or rinse water to the bowel preparation regimen helps eliminate bubble-induced FPs [29][30].

Before the FPs can be effectively reduced, proper training of the endoscopist to recognize and ignore FPs is needed to enable the widespread adoption of the CADe for the detection of colon neoplasms [31].

The optimization of the condition of the bowel lumen can be controlled by the colonoscopist using water exchange colonoscopy, which will be discussed in detail below.

#### 4. Water Exchange and Its Potential Beneficial Effect on Reducing FPs

Among the Gastrointestinal (GI) Endoscopy Editorial Board's top 10 topics in endoscopy in 2019, water exchange (WE) and artificial intelligence (i.e., CADe) were both considered important advances in GI endoscopy [32]. The coincidence brought both to the forefront of the discussion on the improvement of ADR.

Compared with traditional gas (i.e., air or CO<sub>2</sub>) insufflation for colonoscopes, WE is an effective insertion method that minimizes insertion pain and enhances ADR [33][34][35]. It features infusing water to guide the scope advancement in an airless lumen, while suctioning the infused water at the same time during insertion, thus aiming at the almost complete removal of the infused water when cecal intubation is achieved. A network meta-analysis concluded that WE produced the highest ADR when compared with water immersion and gas insufflation [36]. A modified Delphi review also endorsed WE as having better bowel cleanliness, as well as less insertion pain and higher ADRs, than gas insufflation [37].

WE can effectively salvage-clean bubbles and fecal debris during insertion, resulting in better bowel cleanliness during withdrawal. WE consistently showed better Boston Bowel Preparation Scale (BBPS) scores than air insufflation, both in the whole colon and the right colon, the latter of which was usually the dirtiest colon segment [33][34][35]. WE might also help reduce FPs associated with crumpled folds, as there is less need for suction cleaning, and thus the related spasms, during withdrawal [38]. In an analysis of the CADe-overlaid withdrawal phase videos of colonoscopies from an RCT comparing right colon ADR inserted with WE or air insufflation, Tang et al.

WE and CADe both increase ADR but through different mechanisms. WE increases ADRs mainly through insertion salvage cleaning, thus revealing otherwise unexposed polyps (**Table 3**). On the other hand, CADe works as a second observer and points out polyps that are exposed but not recognized due to human error [18]. In other words, the individual strengths of WE and CADe complement the weakness of one another.

**Table 3.** BBPS scores in key randomized controlled trials comparing ADRs between WE and air insufflation.

Study	Sample Size, Air Insufflation vs. WE (n)	Primary Outcome: ADR (95%CI)	Overall BBPS Scores or	Right Colon BBPS Score
Jia et al. [35]	1650 vs. 1653	13.4% vs. 18.3%; RR 1.45 (1.20-1.75) *	7.0 ± 2.3 vs. 7.3 ± 1.6 <sup>#</sup> (Mean ± SD)	2.3 ± 0.7 vs. 2.2 ± 1.5 <sup>#</sup>

Study	Sample Size, Air Insufflation vs. WE (n)	Primary Outcome: ADR (95%CI)	Overall BBPS Scores or	Right Colon BBPS Score
Hsieh et al. [34]	217 vs. 217	37.5% (31.6–44.4%) vs. 49.8% (43–56.4%) *	6.2 ± 1.1 vs. 7.1 ± 1.3 <sup>#</sup> (Mean ± SD)	NA
Cadoni et al. [33]	408 vs. 408	43.4% (35.6–45.3 %) vs. 49.3% (44.3 –54.2 %) *	8.0 (6.0–9.0) vs. 9.0 (7.0–9.0) <sup>#</sup> [Median (IQR)]	2.0 (2.0–3.0) vs. 3.0 (2.0–3.0) <sup>#</sup>

ADR, adenoma detection rate; BBPS, Boston Bowel Preparation Scale; CI, confidence interval; IQR, interquartile range; RR, relative risk; SD, standard deviation; WE: water exchange; \*, p < 0.05; #, p < 0.001; NA, not analyzed.

## References

- Le Clercq, C.M.; Bouwens, M.W.; Rondagh, E.J.; Bakker, C.M.; Keulen, E.T.; de Ridder, R.J.; Winkens, B.; Masclee, A.A.; Sanduleanu, S. Postcolonoscopy colorectal cancers are preventable: A population-based study. *Gut* 2014, 63, 957–963.
- Rex, D.K.; Schoenfeld, P.S.; Cohen, J.; Pike, I.M.; Adler, D.G.; Fennerty, M.B.; Lieb, J.G., 2nd; Park, W.G.; Rizk, M.K.; Sawhney, M.S.; et al. Quality indicators for colonoscopy. *Gastrointest. Endosc.* 2015, 81, 31–53.
- Vinsard, D.G.; Mori, Y.; Misawa, M.; Kudo, S.E.; Rastogi, A.; Bagci, U.; Rex, D.K.; Wallace, M.B. Quality assurance of computer-aided detection and diagnosis in colonoscopy. *Gastrointest. Endosc.* 2019, 90, 55–63.
- Liu, W.N.; Zhang, Y.Y.; Bian, X.Q.; Wang, L.J.; Yang, Q.; Zhang, X.D.; Huang, J. Study on detection rate of polyps and adenomas in artificial-intelligence-aided colonoscopy. *Saudi J. Gastroenterol.* 2020, 26, 13–19.
- Repici, A.; Badalamenti, M.; Maselli, R.; Correale, L.; Radaelli, F.; Rondonotti, E.; Ferrara, E.; Spadaccini, M.; Alkandari, A.; Fugazza, A.; et al. Efficacy of Real-Time Computer-Aided Detection of Colorectal Neoplasia in a Randomized Trial. *Gastroenterology* 2020, 159, 512–520.e7.
- Su, J.R.; Li, Z.; Shao, X.J.; Ji, C.R.; Ji, R.; Zhou, R.C.; Li, G.C.; Liu, G.Q.; He, Y.S.; Zuo, X.L.; et al. Impact of a real-time automatic quality control system on colorectal polyp and adenoma detection: A prospective randomized controlled study (with videos). *Gastrointest. Endosc.* 2020, 91, 415–424.e4.
- Wang, P.; Berzin, T.M.; Glissen Brown, J.R.; Bharadwaj, S.; Becq, A.; Xiao, X.; Liu, P.; Li, L.; Song, Y.; Zhang, D.; et al. Real-time automatic detection system increases colonoscopic polyp and adenoma detection rates: A prospective randomised controlled study. *Gut* 2019, 68, 1813–1819.
- Wang, P.; Liu, X.; Berzin, T.M.; Glissen Brown, J.R.; Liu, P.; Zhou, C.; Lei, L.; Li, L.; Guo, Z.; Lei, S.; et al. Effect of a deep-learning computer-aided detection system on adenoma detection during colonoscopy (CADE-DB trial): A double-blind randomised study. *Lancet Gastroenterol. Hepatol.* 2020, 5, 343–351.
- Hassan, C.; Spadaccini, M.; Iannone, A.; Maselli, R.; Jovani, M.; Chandrasekar, V.T.; Antonelli, G.; Yu, H.; Areia, M.; Dinis-Ribeiro, M.; et al. Performance of artificial intelligence in colonoscopy for adenoma and polyp detection: A systematic review and meta-analysis. *Gastrointest. Endosc.* 2021, 93, 77–85.e6.
- Ahmad, O.F.; Mori, Y.; Misawa, M.; Kudo, S.E.; Anderson, J.T.; Bernal, J.; Berzin, T.M.; Bisschops, R.; Byrne, M.F.; Chen, P.J.; et al. Establishing key research questions for the implementation of artificial intelligence in colonoscopy: A modified Delphi method. *Endoscopy* 2020.
- Abu Dayyeh, B.K.; Thosani, N.; Konda, V.; Wallace, M.B.; Rex, D.K.; Chauhan, S.S.; Hwang, J.H.; Komanduri, S.; Manfredi, M.; Maple, J.T.; et al. ASGE Technology Committee systematic review and meta-analysis assessing the ASGE PIVI thresholds for adopting real-time endoscopic assessment of the histology of diminutive colorectal polyps. *Gastrointest. Endosc.* 2015, 81, 502.e1–502.e16.
- Liu, P.; Wang, P.; Glissen Brown, J.R.; Berzin, T.M.; Zhou, G.; Liu, W.; Xiao, X.; Chen, Z.; Zhang, Z.; Zhou, C.; et al. The single-monitor trial: An embedded CAde system increased adenoma detection during colonoscopy: A prospective randomized study. *Therap. Adv. Gastroenterol.* 2020, 13.
- Mori, Y.; Kudo, S.E.; Misawa, M.; Mori, K. Simultaneous detection and characterization of diminutive polyps with the use of artificial intelligence during colonoscopy. *VideoGIE* 2019, 4, 7–10.
- Ozawa, T.; Ishihara, S.; Fujishiro, M.; Kumagai, Y.; Shichijo, S.; Tada, T. Automated endoscopic detection and classification of colorectal polyps using convolutional neural networks. *Therap. Adv. Gastroenterol.* 2020, 13.
- Lavine, R.A.; Sibert, J.L.; Gokturk, M.; Dickens, B. Eye-tracking measures and human performance in a vigilance task. *Aviat. Space Environ. Med.* 2002, 73, 367–372.
- Chen, Y.; Fu, R.; Xu, Q.; Yuan, W. Mobile Phone Use in a Car-Following Situation: Impact on Time Headway and Effectiveness of Driver's Rear-End Risk Compensation Behavior via a Driving Simulator Study. *Int. J. Environ. Res. Public Health* 2020, 17, 1328.
- Wadhwa, V.; Alagappan, M.; Gonzalez, A.; Gupta, K.; Brown, J.R.G.; Cohen, J.; Sawhney, M.; Pleskow, D.; Berzin, T.M. Physician sentiment toward artificial intelligence (AI) in colonoscopic practice: A survey of US gastroenterologists.

Endosc. Int. Open 2020, 8, E1379–E1384.

18. Misawa, M.; Kudo, S.E.; Mori, Y.; Cho, T.; Kataoka, S.; Yamauchi, A.; Ogawa, Y.; Maeda, Y.; Takeda, K.; Ichimasa, K.; et al. Artificial Intelligence-Assisted Polyp Detection for Colonoscopy: Initial Experience. *Gastroenterology* 2018, 154, 2027–2029.e3.
19. Urban, G.; Tripathi, P.; Alkayali, T.; Mittal, M.; Jalali, F.; Karnes, W.; Baldi, P. Deep Learning Localizes and Identifies Polyps in Real Time With 96% Accuracy in Screening Colonoscopy. *Gastroenterology* 2018, 155, 1069–1078.e8.
20. Becq, A.; Chandnani, M.; Bharadwaj, S.; Baran, B.; Ernest-Suarez, K.; Gabr, M.; Glissen-Brown, J.; Sawhney, M.; Pleskow, D.K.; Berzin, T.M. Effectiveness of a Deep-learning Polyp Detection System in Prospectively Collected Colonoscopy Videos With Variable Bowel Preparation Quality. *J. Clin. Gastroenterol.* 2020, 54, 554–557.
21. Guo, Z.; Nemoto, D.; Zhu, X.; Li, Q.; Aizawa, M.; Utano, K.; Isohata, N.; Endo, S.; Kawarai Lefor, A.; Togashi, K. Polyp detection algorithm can detect small polyps: Ex vivo reading test compared with endoscopists. *Dig. Endosc.* 2021, 33, 162–169.
22. Wang, P.; Xiao, X.; Glissen Brown, J.R.; Berzin, T.M.; Tu, M.; Xiong, F.; Hu, X.; Liu, P.; Song, Y.; Zhang, D.; et al. Development and validation of a deep-learning algorithm for the detection of polyps during colonoscopy. *Nat. Biomed. Eng.* 2018, 2, 741–748.
23. Misawa, M.; Kudo, S.E.; Mori, Y.; Hotta, K.; Ohtsuka, K.; Matsuda, T.; Saito, S.; Kudo, T.; Baba, T.; Ishida, F.; et al. Development of a computer-aided detection system for colonoscopy and a publicly accessible large colonoscopy video database (with video). *Gastrointest. Endosc.* 2021, 93, 960–967.e3.
24. Hassan, C.; Wallace, M.B.; Sharma, P.; Maselli, R.; Craviotto, V.; Spadaccini, M.; Repici, A. New artificial intelligence system: First validation study versus experienced endoscopists for colorectal polyp detection. *Gut* 2020, 69, 799–800.
25. Lee, J.Y.; Jeong, J.; Song, E.M.; Ha, C.; Lee, H.J.; Koo, J.E.; Yang, D.H.; Kim, N.; Byeon, J.S. Real-time detection of colon polyps during colonoscopy using deep learning: Systematic validation with four independent datasets. *Sci. Rep.* 2020, 10, 8379.
26. Podlasek, J.; Heesch, M.; Podlasek, R.; Kilinski, W.; Filip, R. Real-time deep learning-based colorectal polyp localization on clinical video footage achievable with a wide array of hardware configurations. *Endosc. Int. Open* 2021, 9, E741–E748.
27. Holzwanger, E.A.; Bilal, M.; Glissen Brown, J.R.; Singh, S.; Becq, A.; Ernest-Suarez, K.; Berzin, T.M. Benchmarking definitions of false-positive alerts during computer-aided polyp detection in colonoscopy. *Endoscopy* 2020.
28. Chaptini, L.A.; Janec, E.M.; Seltzer, G.; Peikin, S.; Elfant, A.B. Sublingual hyoscine spray as premedication for colonoscopy: A randomized double-blinded placebo-controlled trial. *Am. J. Surg.* 2008, 196, 51–55.
29. Kutyla, M.; O'Connor, S.; Gurusamy, S.R.; Gururatsakul, M.; Gould, K.; Whaley, A.; Kendall, B.J.; Hourigan, L.; Holtmann, G.J. Influence of Simethicone Added to the Rinse Water during Colonoscopies on Polyp Detection Rates: Results of an Unintended Cohort Study. *Digestion* 2018, 98, 217–221.
30. Zhang, S.; Zheng, D.; Wang, J.; Wu, J.; Lei, P.; Luo, Q.; Wang, L.; Zhang, B.; Wang, H.; Cui, Y.; et al. Simethicone improves bowel cleansing with low-volume polyethylene glycol: A multicenter randomized trial. *Endoscopy* 2018, 50, 412–422.
31. Hassan, C.; Badalamenti, M.; Maselli, R.; Correale, L.; Iannone, A.; Radaelli, F.; Rondonotti, E.; Ferrara, E.; Spadaccini, M.; Alkandari, A.; et al. Computer-aided detection-assisted colonoscopy: Classification and relevance of false positives. *Gastrointest. Endosc.* 2020, 92, 900–904.e4.
32. Hwang, J.H.; Jamidar, P.; Kyanam Kabir Baig, K.R.; Leung, F.W.; Lightdale, J.R.; Maranki, J.L.; Okolo, P.I.; 3rd; Swanstrom, L.L.; Chak, A. GIE Editorial Board top 10 topics: Advances in GI endoscopy in 2019. *Gastrointest. Endosc.* 2020, 92, 241–251.
33. Cadoni, S.; Falt, P.; Rondonotti, E.; Radaelli, F.; Fojtik, P.; Gallittu, P.; Liggi, M.; Amato, A.; Paggi, S.; Smajstrla, V.; et al. Water exchange for screening colonoscopy increases adenoma detection rate: A multicenter, double-blinded, randomized controlled trial. *Endoscopy* 2017, 49, 456–467.
34. Hsieh, Y.H.; Tseng, C.W.; Hu, C.T.; Koo, M.; Leung, F.W. Prospective multicenter randomized controlled trial comparing adenoma detection rate in colonoscopy using water exchange, water immersion, and air insufflation. *Gastrointest. Endosc.* 2017, 86, 192–201.
35. Jia, H.; Pan, Y.; Guo, X.; Zhao, L.; Wang, X.; Zhang, L.; Dong, T.; Luo, H.; Ge, Z.; Liu, J.; et al. Water Exchange Method Significantly Improves Adenoma Detection Rate: A Multicenter, Randomized Controlled Trial. *Am. J. Gastroenterol.* 2017, 112, 568–576.
36. Fuccio, L.; Frazzoni, L.; Hassan, C.; La Marca, M.; Paci, V.; Smania, V.; De Bortoli, N.; Bazzoli, F.; Repici, A.; Rex, D.; et al. Water exchange colonoscopy increases adenoma detection rate: A systematic review with network meta-analysis of randomized controlled studies. *Gastrointest. Endosc.* 2018, 88, 589–597.e11.
37. Cadoni, S.; Ishaq, S.; Hassan, C.; Falt, P.; Fuccio, L.; Siau, K.; Leung, J.W.; Anderson, J.; Binmoeller, K.F.; Radaelli, F.; et al. Water-assisted colonoscopy: An international modified Delphi review on definitions and practice recommendations. *Gastrointest. Endosc.* 2020, 93, 1411–1420.e18.
38. Hsieh, Y.H.; Koo, M.; Tseng, C.W.; Yang, H.W.; Leung, F.W. Reduction of multitasking distractions underlies the higher adenoma detection rate of water exchange compared to air insufflation—Blinded analysis of withdrawal phase videos.

## **Keywords**

---

artificial intelligence;computer-aided detection;colonoscopy>false positive;water exchange

---

Retrieved from <https://encyclopedia.pub/12347>