Wearable User Interface

Subjects: Engineering, Mechanical

Contributor: Jung-Hwan Youn , Seung Mo Jeong , Geonwoo Hwang , Hyunwoo Kim , Kyujin Hyeon , Jihwan Park , Ki-Uk Kyung

Making a wearable robot is one of the most important research issues, not only in the bio-mimicking robotics area but also in the conventional robotics or haptics area. There are many advantages for using DEAs for wearable interfaces, such as their small size, flexibility, softness, and customizability. Many wearable haptic interfaces using Dielectric Elastomer Actuators such as tactile displays or vibrotactile displays were suggested from the previous researches. However, there are several challenges in the use of DEAs as actuators embedded into wearable devices such as high driving voltage and low output force compare to the motor based actuator.

Wearable user interface Haptics Dielectric Elastomer Actuators

1. Introduction

Making a wearable robot is one of the most important research issues, not only in the bio-mimicking robotics area, but also in the conventional robotics or haptics area [^[1]]. There are four guidelines to consider when designing a wearable interface: form factor, weight, impairment, and comfort [^[2]]. The wearable interface should be small, compact, and lightweight compared to the body part where the interface will be attached. Also, the wearable interface should not interfere with the wearer's movement, while maintaining comfort. Thus, there are many advantages for using DEAs for wearable interfaces, such as their small size, flexibility, softness, and customizability [^{[3][4][5][6][7][8][9][10][11][12][13][14][15]].}

Portable force feedback devices for surgical simulators have been developed using DEAs. The first portable force feedback device was suggested by Zhang et al. in 2004 [^[3]]. Chain-like actuators made from planar DEAs are attached to the backside of the index finger to give force feedback. They are made of VHB 4910 pre-stretched to 0.109 mm thickness as dielectric elastomer films and a 250 nm thickness of beaten gold as the compliant electrode. The device performs a maximum force of 0.7 N at 4.5 kV under an initial pre-strain force of 4 N, and a maximum elongation of 10.2% at 4.5 kV under a load of 230 g. To enhance the output force of DEAs, multi-layered, multi-element, or multi-layer-element configurations were suggested. One way to improve output force is to use rolled DEAs [^[4]]. The rolled DEAs have a compact structure and enhanced energy density, compared to planar DEAs. They are composed of 30–40 layers of VHB 4910, are 12 mm in diameter, and 45 mm in height. The portable force feedback device using rolled DEAs exerts a maximum force of 7.2 N and displays a maximum elongation of 11% at 3.5 kV.

2. Wearable User Interface using Dielectric Elastomer Actuator

Tactile displays using DEAs have been researched for many years. By using DEAs, the disadvantages of conventional tactile displays, such as difficulty in miniaturization and fabrication, low power density, high cost, and rigidity, can be overcome [^[5]]. The first tactile display using DEAs was suggested by Choi et al. in 2008 [^[5]]. Their system is composed of 4×5 arrays of actuators as shown in Figure 1b. To be completely soft, each actuator is composed of a polymer, especially Nusil CF-2186, with the carbon powder solution as a compliant electrode, a 2 mm diameter and 0.002 g weight. The active layer consists of eight 200 thick layers. The actuator demonstrated an output force of over 13.6 mN at 3.5 kV and 0.1 Hz, which is equivalent to a force-weight ratio of 6.8 N/g.

Pre-straining DE materials could be one solution in enhancing the energy density of DEAs. However, a rigid frame is needed to maintain the pre-strain. To solve this issue, a method of embedding nylon fibers into the DEA was suggested [^[6]]. The insertion of nylon does not deteriorate the breakdown voltage nor operation of a DEA due to the nylon's insulation property and low bending stiffness. A nylon-embedded module is attached to the forearm and conveys the shear force to the wearer. As another structure of DE tactile display, hydrostatically-coupled DEAs (HC-DEA) were suggested by Carpi et al. in 2010 [[2]]. The configuration of HC-DEAs for a tactile interface is shown in Figure 1c. The bubble-like form of HC-DEA is due to silicone-oil placed between the active membrane and the passive membrane. Since the active membrane is electrically connected and the passive membrane is the end-effector, the silicone oil's placement ensures electrical safety [^[8]]. Both active and passive membranes consist of bi-axially 400% pre-stretched VHB 4910. Carbon grease is used as a compliant electrode to the active membrane. This actuator is 10 mm in radius and 7 mm in height. In this research, the maximum force is 0.6 N at a working voltage of 4.5 kV. To improve the output force of the HC-DEA, adopting a stacked active membrane was introduced ^[9]]. The active membrane is made of three layers of bi-axially 300% pre-stretched VHB 4910 and carbon grease as a compliant electrode. The thickness of the passive membrane is also matched to maintain the shape of the actuator. As a result, the maximum force is measured as 0.7 N under the working voltage of 4.0 kV. By using stacked HC-DEAs and increasing the thickness of the passive membrane, not only was the output force enhanced, but the working voltage was also decreased.



Figure 1. Wearable interfaces using DEAs: (**a**) force feedback system using roll DEA [120]; (**b**) soft tactile display using DEA [121]; (**c**) tactile display using HC-DEAs [124]; (**d**) stacked DEA with polymeric frame to maintain curvature [126].

Lower sustainability against pressurized contact is also a limitation of DEAs used as a wearable interface [10]. As shown in Figure 1d, a perforated protecting polymer is formed on the actuator to prevent the actuator from being initially pressurized. This structure can be utilized to create a glove-type tactile display, that can give embossed feedback at fingertips, or a patched display that can give vibro-tactile feedback to the forearm. This actuator consists of six stacks of 100 Elastosil P7670 as dielectric elastomers, and silver nanowires (AgNWs) as compliant electrodes. The diameter of the active area is 15 mm. The maximum force is reported as 255 mN at 4 kV and the resonant frequency is 191 Hz, while the maximum vertical protrusion is 650 μ m under an unloading condition and at 4 kV and 1 Hz.

Research about the development of DE micro-actuators, which concentrate force on a localized area, was introduced in 2020 [^{11]}]. The DE micro-actuators are fabricated through photolithography, and the dimension of each micro-actuator is 200 μ m × 7 mm × 400 μ m. They are made of Sylgard 184 as the dielectric elastomer and AgNWs as compliant electrodes. With these DE micro-actuators, the tactile display is organized into an array, generating more than 30 times the force of the forearm vibro-tactile displacement perception threshold at 20 MV/m.

There are also other possible wearable interface applications of DEAs. An active compression bandage using DEAs was suggested. It can be taken on and off easily, fits different calf sizes, and is provided in small size modules [12]. Also, the compression magnitude of the bandage can be modulated by changing the initial prestretch ratio or working voltage. A hand tremor suppression using DEAs was also introduced [13]. Undesired motions by Parkinsonian tremors have faster frequency (more than 3 Hz) than voluntary motion (0~2 Hz). By creating the motion which has that frequency of tremor, the movement can be canceled out. DEAs are attached to the wrist to reduce tremors in the flexion and extension direction. In addition to presenting these possibilities, researches on controller design and parameter design are being conducted [^{14]},^[15]]. Recently, research applying textile DEAs on a wearable interface has been conducted. In particular, replacing compliant electrodes with stretchable and conductive textiles have been investigated [^{16]}].

Wearable interfaces represent one of the recently emerging applications of DEAs. However, there are several challenges in the use of DEAs as actuators embedded into wearable devices. For example, reducing driving voltage is an issue to be solved with a high priority. High operating voltage often causes additional cost in terms of electronics as well as insulation. In addition, the insulation layers lead to output force degradation. Therefore, for low driving voltage DEA devices, we need to consider various factors including structural design, electronics cost, insulation material, as well as new DEA material with a high dielectric constant. In order to use DEA for haptic interfaces, resonant frequency needs to be considered for the efficient delivery of tactile cues to users. For example, human indentation sense is sensitive at a low frequency around 1~10 Hz, and vibro-tactile sense is the most sensitive around 200 Hz [10]. So, to this purpose, we can consider a design with a specific frequency range to maximize perceived tactile sense.

References

- 1. Guoying Gu; Jian Zhu; Li-Min Zhu; Xiangyang Zhu; A survey on dielectric elastomer actuators for soft robots. *Bioinspiration & Biomimetics* **2017**, *12*, 11003, 10.1088/1748-3190/12/1/011003.
- Claudio Pacchierotti; Stephen Sinclair; Massimiliano Solazzi; Antonio Frisoli; Vincent Hayward; Domenico Prattichizzo; Wearable Haptic Systems for the Fingertip and the Hand: Taxonomy, Review, and Perspectives. *IEEE Transactions on Haptics* 2017, *10*, 580-600, 10.1109/toh.2017.2 689006.
- Zhang, R.; Kunz, A.; Gabor, K.; Silvain, M.; Mazzon, A. Dielectric Elastomer Actuators for a Portable Force Feedback Device. In Proceedings of the Eurohaptics 04, Muenchen, Germany, 5– 7 June 2004; pp. 300–307.
- Zhang, R.; Kunz, A.; Lochmatter, P.; Kovacs, G. Dielectric Elastomer Spring Roll Actuators for a Portable Force Feedback Device. In Proceedings of the 2006 14th Symposium on Haptic Interfaces for Virtual Environment and Teleoperator Systems, Arlington, VA, USA, 25–26 March 2006; pp. 347–353.
- 5. Mohammad Saadeh; Mohamed Trabia; Yantao Shen; Sami Fadali; Design of a Wearable Fingertip Haptic Braille Device. *Journal of Medical Devices* **2010**, *4*, 027524, 10.1115/1.3443233.
- Christian Bolzmacher; James Biggs; Mandayam Srinivasan; Flexible dielectric elastomer actuators for wearable human-machine interfaces. SPIE Smart Structures and Materials + Nondestructive Evaluation and Health Monitoring 2006, 6168, 616804-616804, 10.1117/12.67798
 1.

- 7. Federico Carpi; Gabriele Frediani; Danilo De Rossi; Hydrostatically coupled dielectric elastomer actuators for tactile displays and cutaneous stimulators. *SPIE Smart Structures and Materials* + *Nondestructive Evaluation and Health Monitoring* **2010**, *7642*, 76420, 10.1117/12.847562.
- Gabriele Frediani; Daniele Mazzei; Danilo Emilio De Rossi; Federico Carpi; Wearable Wireless Tactile Display for Virtual Interactions with Soft Bodies. *Frontiers in Bioengineering and Biotechnology* 2014, 2, 10.3389/fbioe.2014.00031.
- Hugh Boys; Gabriele Frediani; Stefan Poslad; James Busfield; Federico Carpi; A dielectric elastomer actuator-based tactile display for multiple fingertip interaction with virtual soft bodies. *Electroactive Polymer Actuators and Devices (EAPAD) 2017* 2017, 10163, 101632, 10.1117/12.2 259957.
- Seongcheol Mun; Sungryul Yun; SaeKwang Nam; Seung Koo Park; Suntak Park; Bong Park; Jeong Mook Lim; Ki-Uk Kyung; Electro-Active Polymer Based Soft Tactile Interface for Wearable Devices. *IEEE Transactions on Haptics* **2018**, *11*, 15-21, 10.1109/toh.2018.2805901.
- 11. Sungryul Yun; Suntak Park; Bongje Park; Semin Ryu; Sm Jeong; Ki-Uk Kyung; A Soft and Transparent Visuo-Haptic Interface Pursuing Wearable Devices. *IEEE Transactions on Industrial Electronics* **2020**, *67*, 717-724, 10.1109/tie.2019.2898620.
- 12. S Pourazadi; S Ahmadi; C Menon; Towards the development of active compression bandages using dielectric elastomer actuators. *Smart Materials and Structures* **2014**, *23*, 65007, 10.1088/09 64-1726/23/6/065007.
- Christopher R. Kelley; Jeffrey L. Kauffman; Exploring dielectric elastomers as actuators for hand tremor suppression. *Electroactive Polymer Actuators and Devices (EAPAD) 2017* 2017, 10163, 1016322, 10.1117/12.2258735.
- Christopher R Kelley; Jeffrey L. Kauffman; Tremor suppression controller for dielectric elastomer orthosis. *Electroactive Polymer Actuators and Devices (EAPAD) XX* 2018, 10594, 105940U, 10.1 117/12.2296697.
- Christopher R. Kelley; Jeffrey L. Kauffman; Necessary dielectric elastomer parameters for wearable tremor suppression. *Electroactive Polymer Actuators and Devices (EAPAD) XXI* 2019, 10966, 109661T, 10.1117/12.2514335.
- Guo, J.; Xiang, C.; Helps, T.; Rossiter, J. Electroactive textile actuators for wearable and soft robots. In Proceedings of the Robosoft 2018, Livorno, Italy, 24–28 April 2018.

Retrieved from https://encyclopedia.pub/entry/history/show/8551