Swallow Detection with Acoustics and Accelerometric-Based Wearable Technology

Subjects: Otorhinolaryngology | Primary Health Care | Engineering, Biomedical

Contributor: Bryan Pak-Hei So , Tim Tin-Chun Chan , Liangchao Liu , Calvin Chi-Kong Yip , Hyo-Jung Lim , Wing-Kai Lam , Duo Wai-Chi Wong , Daphne Sze Ki Cheung , James Chung-Wai Cheung

Swallowing disorders, especially dysphagia, might lead to malnutrition and dehydration and could potentially lead to fatal aspiration. Benchmark swallowing assessments, such as videofluoroscopy or endoscopy, are expensive and invasive. Wearable technologies using acoustics and accelerometric sensors could offer opportunities for accessible and home-based long-term assessment. Identifying valid swallow events is the first step before enabling the technology for clinical applications.

dysphagia deglutition disorder eating disorder

1. Introduction

Swallowing is a natural yet essential part of the daily life. Human performs spontaneous swallowing (saliva and food/drink) 0.98 times per minute on average [1]. With different definitions and measurement techniques, Lear et al. ^[2] suggested that humans swallow approximately 200 to 1000 times a day, while Rudney et al. ^[3] reported that spontaneous swallows are performed by healthy humans 18 to 400 times per hour. However, some people may have difficulty swallowing, especially aged people or people with chronic conditions. Swallowing difficulty is also termed dysphagia, in which dysphagic individuals have problems chewing and swallowing food or liquids, experience pain during swallowing, or even be unable to swallow. Notably, the bolus may enter the airway and lungs, leading to aspiration pneumonia, which is fatal but clinically silent [4]. Dysphagia is generally chronic but deteriorates with the worsening of cognition and functions in the progression of dementia or other neurological disorders ^{[5][6]}. Therefore, continuous monitoring or assessment could be necessary to identify the stage at high risk of choking or aspiration for timely management and rehabilitation [7]. In addition, dysphagia patients may be reluctant to eat due to the fear of choking, pain, or difficulty that causes malnutrition, dehydration, depression, and anorexia ^[9]. More than one-third of older adults reported dysphagia or swallowing disorders during their lifetime, which were associated with stroke, diabetes, Parkinson's, and Alzheimer's disease [10][11]. Howden [12] and Ney et al. [13] reported that the prevalence of dysphagia could be 22% and 40% for seniors aged over 50 and 60, respectively. A recent survey reported that swallowing difficulty was reported in one in every six adults, and some of them might not seek medical care $\frac{14}{2}$.

Swallowing assessment or monitoring is imperative to facilitate early diagnosis, management, or rehabilitation to reduce mortality and improve the quality of life for dysphagia individuals. Nowadays, the Videofluoroscopic Swallowing Study (VFSS) and Fiberoptic Endoscopic Evaluation of Swallowing (FEES) are golden standards for

instrumented assessment ^[15]. VFSS applies a dynamic fluoroscopic imaging technique to visualize the detailed swallowing process in oral, pharyngeal, laryngeal, and oesophageal regions in real-time ^[16]. For FEES, practitioners inspect the postural maneuvers of the nasal structures when the patients speak, eat, and breathe using an endoscope ^[17]. However, VFSS and FEES are expensive, cause discomfort and risks to the patients, and can only be conducted occasionally.

Non-instrumental bedside assessments for swallowing are alternatives to compromise cost and test frequency that could be readily adopted in nursing homes or care homes by an occupational therapist or speech therapist. A standard bedside screening process involves anamnesis assessment, morphodynamical evaluation, gustative function with specific stimulation test, and the oral feeding test ^[18]. Other related tests include the 3-ounce water swallowing test ^[19], cough reflex test ^[20], and cervical auscultation, which uses a stethoscope to amplify and listen to the swallowing sound ^[21]. Most of these instruments lacked sensitivity and predictive strength and poor reproducibility and consistency in the protocols ^{[21][22]} but could be routinely conducted for initial screening of swallowing functions ^[23].

Cervical auscultation refers to the measurement of sound or vibration of the throat for swallowing assessment, which is traditionally conducted by physicians using a stethoscope ^[24]. Wearable technology, such as accelerometry, acoustics, and electromyogram, could be more robust to facilitate non-invasive and non-ionizing, continuous monitoring or screening with less cost. Swallowing accelerometry monitors the translation of vibration through the aerodigestive tract and hyoid bone kinetics during swallowing ^[25]. The acoustic technique uses an inexpensive microphone to record swallowing sounds and may sometimes integrate with the accelerometry approach ^[26]. Takahashi et al. ^[27] could be among the pioneers that systematically reviewed and evaluated acoustic methods for the detection of swallowing sounds, while Taveira et al. ^[28] reviewed and compared the diagnostic validity of swallowing-sound-based methods to videofluoroscopy. Thereafter, more developments have been conducted using multimodal sensors, advanced data processing techniques and machine learning models.

2. Instrument Configuration

Among the 11 eligible articles, five of them utilized only acoustics (microphone) ^{[29][30][31][32][33]}, one utilized only accelerometers in the instrument ^[34], and five applied a multimodal system ^{[35][36][37][38][39]}. However, two articles on multimodal systems did not fully describe the modalities other than acoustics ^{[35][38]}. Other multimodal systems involved surface electromyography (sEMG), mechanomyography (MMG), and airflow pressure sensor.

As shown in **Table 1**, a single microphone for detecting swallowing sounds appeared in three articles ^{[29][32][33]}. Skowronski et al. ^[33] made use of a miniature surface-mounted microphone and characterized the signal using Human Factor Cepstral Coefficients ^[40], which was originally used for automatic speech recognition. Bi et al. ^[29] developed the "AutoDietary" system using a throat microphone. The system also displayed the food type recognition results for the users for personal health management. Kurihara et al. ^[32] customized the device by attaching a bi-directional electret condenser microphone on the ends of an air tube to detect the swallowing microphone through the pressure propagation along the air tube. Two studies employed two microphones but with

different principles ^{[30][31]}. The major laryngeal microphone was used to record the swallowing sound directly in both cases. On the one hand, Fukuike et al. ^[31] further improved the system accuracy by adding a condenser microphone on the nostril. On the other hand, Fontana et al. ^[30] used the condenser microphone to detect the swallowing sound in the subsonic range. Additionally, Amft and Troster ^[36] integrated a stethoscope microphone with sEMG of the cricopharyngeus muscle to recognize swallowing. They also presented separate analyses on dietary movement activity and chewing activity recognition using other sensors ^[36].

Author (Year)	Sensors	Location	Swallowing or Non-swallowing events	
	Miniature ACC (NM)	Level of thyroid cartilage		
Afkari ^[35]	sEMG (NM)	sEMG (NM) Level of cricopharyngeus muscle mL of water as fa	dry (saliva) swallowing, drink 100 mL of water as fast as possible.	
	Omnidirectional electret MIC (NM)	Level of cricoid cartilage		
Amft and	sEMG (Nexus-10, MindMedia)	Collar at infra-hyoid throat region	Participants were allowed to move, chew, & speak normally during the recording. The participants were	
Troster [<u>36</u>]	Stethoscope MIC (ECM- C115, Sony)	Collar below hyoid	asked to drink 5 mL & 15 mL of water, eat a spoonful of yogurt, & 2 cm ³ of bread in one piece.	
Bi et al. ^[29]	Throat MIC [NM]	Over neck close to the jaw	Apple, carrot, chip, cookie, peanut, walnut, water.	
Fontana et	Condenser MIC (CZN-15E)	thyroid cartilage level, one side of the neck	Start with 5 min quiet sitting, 5 min reading aloud, a meal of 4 food	
al. ^[30]	Piezoelectric MIC (IASUS NT, IASUS Concept Ltd.)	Over laryngopharynx	fact yogurt, 250 mL water) was consumed at unlimited time.	
Fukuike et	Condenser MIC (WM-61A, Panasonic, Osaka, Japan)	Fixed on a silicone tube and placed inside the left nostril	Taking a meal and stepping on a foot pedal when swallowed.	
а.	Laryngeal MIC (SH-12iK, Nanzu, Shizuoka, Japan)	Over anterior larynx	rawn, cougn, sign, throat clearing, gargling, and sipping tea.	
Kurihara et al. ^[32]	Bi-directional electret condenser MIC (EM114, Primo Co., Ltd.)	MIC attached to air tube hung over neck with anterior opening	swallowing nothing, tea (10 mL), tea with a thickener (10 mL), rice cake (10 g).	
Lee et al. [<u>37</u>]	Dual axis ACC (ADXL322)	Below thyroid cartilage aligned in anterior-	Water, barium suspension (Ba), nectar-thick apple juice (Ne),	

Table 1. Instrument setting, location, and assessment procedures in the reviewed articles.

Submental mechanomyography (developed by Silva and Chau ⁽⁴¹⁾) On the geniohyoid On the geniohyoid honey-thick apple juice (Ho), spon-thick apple juice (Sp). Pressure Transducer (PTAFLITE, Glass Technologies) At nasal cannula honey-thick apple juice (Sp). Makeyev et al. ¹³⁹ Throat microphone (IASUS NT, IASUS Concept Ltd.)* At nasal cannula Start with 10 min silent, 10 min reading aloud, meal of mixed size consumed at an unlimited time (including cheese pizza, yogurg, apple, peanut butter sandiwtch), 10 min silent, 10 min reading aloud. Sazonov et al. ¹³⁹ Throat microphone (IASUS NT, IASUS Concept Ltd.)* Over laryngopharynx (ADXL322) 20 min rest, a meal, then 20 min rest. Sejdic et al. Dual-axis accelerometer (ADXL322) Anterior to cricoid cartilage, along anterior- posterior and superior- inferior axes 5 mL liquid, dry swallow, drink water in natural & chin-tucked position atural & chin-tucked position min silent, 10 min, rest, and move, yawn, sniff, tongue move, speech, hum, throat clear, cough.	Author (Year)	Sensors	Location	Swallowing or Non-swallowing events
Submental mechanomyography (developed by Silva and Chau (4))On the geniohyoidPressure Transducer (PTAFLITE, Glass Technologies)At nasal cannulaMakeyev et al. 1840Throat microphone (IASUS NT, IASUS Concept Ltd.)*Over laryngopharynzyStart with 10 min silent, 10 min reading aloud, meal of mixed size consumed at an unlimited time (including cheese pizza, yogurg, apple, peanut butter sandiwtch.) 10 min silent, 10 min reading aloud, meal of mixed size consumed at an unlimited time (including cheese pizza, yogurg, apple, peanut butter sandiwtch.) 10 min silent, 10 min reading aloud, min silent, 10 min reading aloud, mi			posterior and superior- inferior axes	honey-thick apple juice (Ho), spoon-thick apple juice (Sp).
Pressure Transducer (PTAFLITE, Glass Technologies)At nasal cannulaMakeyev et al. 1989Throat microphone (IASUS NT, IASUS Concept Ltd.)*Start with 10 min silent, 10 min reading aloud, meal of mixed size consumed at an unlimited time 		Submental mechanomyography (developed by Silva and Chau ^[41])	On the geniohyoid	
Makeyev et al. 1990Throat microphone (IASUS NT, IASUS Concept Ltd.)*Over laryngopharynxStart with 10 min silent, 10 min reading aloud, meal of mixed size consumed at an unlimited time (including cheese pizza, yogurg, aple, peanut butter sandiwtch), 10 min silent, 10 min reading aloud.Sazonov et al. 1990Throat microphone (IASUS NT, IASUS Concept Ltd.)*Over laryngopharynx20 min rest, a meal, then 20 min 		Pressure Transducer (PTAFLITE, Glass Technologies)	At nasal cannula	
Sazonov et al. [39]Throat microphone (IASUS NT, IASUS Concept Ltd.)*Over laryngopharynx20 min rest, a meal, then 20 min rest.Sejdic et al. [34]Dual-axis accelerometer (ADXL322)Anterior to cricoid cartilage, along anterior- 	Makeyev et al. ^[<u>38</u>]	Throat microphone (IASUS NT, IASUS Concept Ltd.) *	Over laryngopharynx	Start with 10 min silent, 10 min reading aloud, meal of mixed size consumed at an unlimited time (including cheese pizza, yogurg, apple, peanut butter sandiwtch), 10 min silent, 10 min reading aloud.
Sejdic et al. [34]Dual-axis accelerometer (ADXL322)Anterior to cricoid cartilage, along anterior- posterior and superior- inferior axes [35]dry (saliva) swallow, drink water in natural & chin-tucked position34[[37] Skowronski et al. [33]35Miniature surface-mounted MIC (VT506, Voice Technologies, Zurich, [37] Switzerland)Laterally below the cricoid cartilage (ADXL322)5 mL liquid, dry swallow, head move, yawn, sniff, tongue move, speech, hum, throat clear, cough.	Sazonov et al. ^[39]	Throat microphone (IASUS NT, IASUS Concept Ltd.) *	Over laryngopharynx	20 min rest, a meal, then 20 min rest.
34][37] Skowronski35]Miniature surface-mounted5 mL liquid, dry swallow, head move, yawn, sniff, tongue move, speech, hum, throat clear, cough.34][37] Switzerland)Laterally below the cricoid cartilage5 mL liquid, dry swallow, head move, yawn, sniff, tongue move, 	Sejdic et al. [<u>34</u>]	Dual-axis accelerometer (ADXL322)	Anterior to cricoid cartilage, along anterior- posterior and superior- inferior axes	dry (saliva) swallow, drink water in natural & chin-tucked position
	[<mark>34][37]</mark> Skowronski et al. ^[33]	[35] ^{Miniature surface-mounted} MIC (VT506, Voice Technologies, Zurich, ^[37] Switzerland) [41]	Laterally below the cricoid cartilage	5 mL liquid, dry swallow, head move, yawn, sniff, tongue move, speech, hum, throat clear, cough.

anterior-posterior and superior-inferior directions [34][35][37].

There were variations in the locations of the sensors, which may depend on the types and the suspension methods. Although few studies vaguely mentioned that the sensors shall be attached over the laryngopharynx, thyroid cartilage and cricoid cartilage were two anatomical landmarks highlighted ^{[33][34][35][37]}. The sensors could be glued or taped to the throat surface ^{[34][35]}, collared ^[36], or in the form of a necklace ^{[30][31][32][37]}.

3. Assessment Protocol for Swallowing

Since swallowing is a continuous process, segmenting a time frame to stamp the swallowing episode is essential to define the "sample counts" for evaluating accuracy. The episode stamping method could be classified as event-based or episode-based. Two studies attempted both event-based and episode-based approaches for the

evaluation [38][39]. For the other studies, five [29][31][32][34][35] adopted the event-based approach, and four [30][33][36][37] adopted the episode-based approach, respectively.

For event-based stamping, the conditions were controlled, and the researchers instructed the participants to perform one maneuver at a time, in which the event could be easily labeled for a period. For the epoch-based approach, the participants were often free to conduct a series of activities at each time. Then, the time was sliced into several non-overlapping time units (epochs) by algorithms or data processing techniques and was then manually labeled by revisiting the videotape. Alternatively, participants might be asked to press a button or pedal during their swallowing process for labeling ^{[30][31]}.

The swallowing protocol could be broadly classified as non-swallowing maneuvers and swallowing maneuvers, while some studies attempted to have a fine-grained classification within these two categories (Table 1). For non-swallowing, the dry swallow was referred to as saliva swallowing ^{[33][34][35]}, while assessing non-swallowing through silence or talking was often implemented through an epoch-based approach (detailed in the next paragraph) ^{[30][36]} ^{[38][39]}. Some studies investigated different types of throat movements as non-swallowing events, including yawning, coughing, sighing, sniffing, throat clearing, gargling, speech, and tongue moving ^{[31][33]}. Besides, it shall be noted that Fukuike et al. ^[31] considered sipping tea as a non-swallowing maneuver. On the other hand, there was no consensus on the kinds of food to prompt swallowing events. For the epoch-based approach, participants were asked to take a meal with a variety of food without controlling participants to eat one kind of food at a time during the data collection. Besides, drinking water appeared in most of the articles ^{[29][30][33][34][35][36][37]}, while yogurt was the most famous semifluid food ^{[30][36][38]}. For solid food, bread, crackers, cookies, pizza, sandwiches, fruit, and peanuts were some examples considered ^{[29][30][36][38]}.

4. Segmentation and Feature Extraction Strategy

Researchers had to identify whether a swallowing event happened within a time frame because of the continuous nature of swallowing. Two studies manually segmented the time window ^{[33][35]}, while four studies specified the duration of the segmented time window, ranging from 200 ms to 1.5 s ^{[30][36][37][38]}. Fukuike et al. ^[31], Kurihara et al. ^[32], and Sejdic et al. ^[34] utilized the semblable wave period, template matching, and minimum description length-based segmentation, respectively. Two studies accounted for randomized sampling concepts in the segmentation process, including the Hidden Markov Model (HMM) conducted by Bi et al. ^[29] and the grid search conducted by Sazonov et al. ^[39].

For the feature extraction strategy, four studies exploited the time-domain raw signals for classification ^{[30][31][34][35]}, while one made use of the frequency-domain raw signals ^[39]. Predetermined features were computed for analysis in three articles ^{[29][32][37]}. For example, Amft and Troster ^[36] considered and fused the spectral features (band energy, autocorrelation coefficient, and energy) and EMG features (total and maximum). Three studies performed some data reduction processes and established specific index parameters before the classification process ^{[33][36]}. ^[38], such as using Principal Component Analysis (PCA).

5. Classification and Performance

Depending on the nature of the classification (i.e., swallowing vs. non-swallowing or classification of different food types) and the stamping approach (i.e., event-based vs. epoch-based), studies might apply different classification approaches. In order to classify/identify the swallowing event, three studies applied a threshold-based approach ^[30] ^{[31][35]}, while others implemented statistical or machine learning models ^{[29][32][33][34][36][37][38][39]}. These models included logistic regression, decision tree, Gaussian Mixture Model (GMM), Support Vector Machine (SVM), Artificial Neural Network (ANN), etc.

For the threshold-based approach, a swallowing event was often recognized whenever the collected signal exceeded a predefined threshold value for more than a certain time. Nevertheless, the cut-off level or time range was not adequately justified in the papers, and most of them were empirical. Amft and Troster ^[36] applied compared acoustics, accelerometry, and EMG data with a set of reference voltages and integrated them by a logic gate (AND) but without justifying the source of the reference set. Fontana et al. ^[30] established individualized threshold levels based on the collected signal during a reading task. They also suggested that the time range threshold shall be 0.6 s ^[30], which was an estimated time for a complete swallow ^[39]. On the other hand, Fukuike et al. ^[31] decided to use twice the mean baseline as the threshold level, and a recognized event shall last longer than 0.35 s.

For the evaluation of classification performance, accuracy, sensitivity, specificity, and positive predictive value (PPV) are common evaluation metrics. Sensitivity and PPV are also sometimes termed precision and recall from the perspective of information retrieval in the field of data science ^[42]. Sensitivity represented the proportion of recognizing a swallowing event/class when that event/class did occur, while specificity was the proportion of recognizing not a swallowing event/class when that event/class had not occurred. Accuracy is the ratio of correct classifications over the total number of tests. Besides, one study ^[29] supplemented the receiver operating characteristics (ROC) curve to demonstrate the discrimination capacity.

As a rule of thumb, classifiers required an independent dataset for training and testing (model evaluation) to better evaluate the generalizing capability. Sejdic et al. ^[34] evaluated the model using both synthetic tests and real swallowing signals. Despite a different number of folds, most of the model-based classifiers applied k-fold cross-validation, while Kurihara et al. ^[32] adopted a leave-one-out approach. In addition, Lee et al. ^[37] calculated the accuracy metrics based on a bootstrapping augmentation after a 10-fold cross-validation of the model to account for the unbalanced class sizes.

The 11 reviewed articles involved 15 classifiers in the data synthesis. There was a high variation in accuracy level among studies, ranging from 68.2% to 96.8%. Some studies had a classification performance as unreliable as a random guess (40–60%). Besides, despite that the accuracy metric of the review articles is generally satisfactory, the outcomes of other metrics (such as sensitivity, specificity, and PPV) could be quite different between studies. For example, Makeyev et al. ^[38] attained 44% sensitivity and 99% specificity in their epoch-based SVM model. Amft and Troster ^[36] got 20% positive predictive value and 68% sensitivity in their classification method using the

agreement of detectors. The reason could be due to the problem of imbalanced class size, especially for epochbased approaches.

References

- Bulmer, J.M.; Ewers, C.; Drinnan, M.J.; Ewan, V.C. Evaluation of Spontaneous Swallow Frequency in Healthy People and Those With, or at Risk of Developing, Dysphagia: A Review. Gerontol. Geriatr. Med. 2021, 7, 23337214211041801.
- 2. Lear, C.S.; Flanagan, J., Jr.; Moorrees, C. The frequency of deglutition in man. Arch. Oral Biol. 1965, 10, 83-IN15.
- Rudney, J.; Ji, Z.; Larson, C. The prediction of saliva swallowing frequency in humans from estimates of salivary flow rate and the volume of saliva swallowed. Arch. Oral Biol. 1995, 40, 507– 512.
- 4. Hammond, C.A.S.; Goldstein, L.B. Cough and aspiration of food and liquids due to oralpharyngeal dysphagia: ACCP evidence-based clinical practice guidelines. Chest 2006, 129, 154S–168S.
- Namasivayam-MacDonald, A.M.; Alomari, N.; Attner, L.; Benjamin, R.D.; Chill, A.; Doka, S.; Guastella, R.; Marchese, J.; Oppedisano, S.; Ressa, K. A retrospective analysis of swallowing function and physiology in patients living with dementia. Dysphagia 2022, 37, 900–908.
- De Stefano, A.; Di Giovanni, P.; Kulamarva, G.; Gennachi, S.; Di Fonzo, F.; Sallustio, V.; Patrocinio, D.; Candido, S.; Lamarca, G.; Dispenza, F. Oropharyngeal dysphagia in elderly population suffering from mild cognitive impairment and mild dementia: Understanding the link. Am. J. Otolaryngol. 2020, 41, 102501.
- Furuya, H.; Kikutani, T.; Igarashi, K.; Sagawa, K.; Yajima, Y.; Machida, R.; Tohara, T.; Takahashi, N.; Tamura, F. Effect of dysphagia rehabilitation in patients receiving enteral nutrition at home nursing care: A retrospective cohort study. J. Oral Rehabil. 2020, 47, 977–982.
- 8. Okuni, I.; Ebihara, S. Are Oropharyngeal Dysphagia Screening Tests Effective in Preventing Pneumonia? J. Clin. Med. 2022, 11, 370.
- Baijens, L.W.; Clavé, P.; Cras, P.; Ekberg, O.; Forster, A.; Kolb, G.F.; Leners, J.-C.; Masiero, S.; Mateos-Nozal, J.; Ortega, O. European Society for Swallowing Disorders–European Union Geriatric Medicine Society white paper: Oropharyngeal dysphagia as a geriatric syndrome. Clin. Interv. Aging 2016, 11, 1403.
- Roy, N.; Stemple, J.; Merrill, R.M.; Thomas, L. Dysphagia in the elderly: Preliminary evidence of prevalence, risk factors, and socioemotional effects. Ann. Otol. Rhinol. Laryngol. 2007, 116, 858–

865.

- Wirth, R.; Dziewas, R.; Beck, A.M.; Clavé, P.; Hamdy, S.; Heppner, H.J.; Langmore, S.; Leischker, A.H.; Martino, R.; Pluschinski, P. Oropharyngeal dysphagia in older persons–from pathophysiology to adequate intervention: A review and summary of an international expert meeting. Clin. Interv. Aging 2016, 11, 189.
- 12. Howden, C.W. Management of acid-related disorders in patients with dysphagia. Am. J. Med. Suppl. 2004, 117, 44–48.
- 13. Ney, D.M.; Weiss, J.M.; Kind, A.J.; Robbins, J. Senescent swallowing: Impact, strategies, and interventions. Nutr. Clin. Pract. 2009, 24, 395–413.
- Adkins, C.; Takakura, W.; Spiegel, B.M.; Lu, M.; Vera-Llonch, M.; Williams, J.; Almario, C.V. Prevalence and characteristics of dysphagia based on a population-based survey. Clin. Gastroenterol. Hepatol. 2020, 18, 1970–1979. e1972.
- 15. Steele, C.M. The blind scientists and the elephant of swallowing: A review of instrumental perspectives on swallowing physiology. J. Texture Stud. 2015, 46, 122–137.
- 16. Boaden, E.; Nightingale, J.; Bradbury, C.; Hives, L.; Georgiou, R. Clinical practice guidelines for videofluoroscopic swallowing studies: A systematic review. Radiography 2020, 26, 154–162.
- Prikladnicki, A.; Santana, M.G.; Cardoso, M.C. Protocols and assessment procedures in fiberoptic endoscopic evaluation of swallowing: An updated systematic review. Braz. J. Otorhinolaryngol. 2021, 88, 445–470.
- 18. Maccarini, A.R.; Filippini, A.; Padovani, D.; Limarzi, M.; Loffredo, M.; Casolino, D. Clinical noninstrumental evaluation of dysphagia. Acta Otorhinolaryngol. Ital. 2007, 27, 299–305.
- 19. Suiter, D.M.; Leder, S.B. Clinical utility of the 3-ounce water swallow test. Dysphagia 2008, 23, 244–250.
- 20. Lee, J.Y.; Kim, D.-K.; Seo, K.M.; Kang, S.H. Usefulness of the simplified cough test in evaluating cough reflex sensitivity as a screening test for silent aspiration. Ann. Rehabil. Med. 2014, 38, 476.
- 21. Lagarde, M.L.; Kamalski, D.M.; Van Den Engel-Hoek, L. The reliability and validity of cervical auscultation in the diagnosis of dysphagia: A systematic review. Clin. Rehabil. 2016, 30, 199–207.
- 22. O'Horo, J.C.; Rogus-Pulia, N.; Garcia-Arguello, L.; Robbins, J.; Safdar, N. Bedside diagnosis of dysphagia: A systematic review. J. Hosp. Med. 2015, 10, 256–265.
- Shem, K.L.; Castillo, K.; Wong, S.L.; Chang, J.; Kao, M.C.; Kolakowsky-Hayner, S.A. Diagnostic accuracy of bedside swallow evaluation versus videofluoroscopy to assess dysphagia in individuals with tetraplegia. PMR 2012, 4, 283–289.

- Dudik, J.M.; Coyle, J.L.; Sejdić, E. Dysphagia screening: Contributions of cervical auscultation signals and modern signal-processing techniques. IEEE Trans. Hum.-Mach. Syst. 2015, 45, 465– 477.
- 25. Zahnd, E.; Movahedi, F.; Coyle, J.L.; Sejdić, E.; Menon, P.G. Correlating tri-accelerometer swallowing vibrations and hyoid bone movement in patients with dysphagia. In Proceedings of the ASME International Mechanical Engineering Congress and Exposition, Phoenix, AZ, USA, 11–17 November 2016; p. V003T004A083.
- 26. Reynolds, E.W.; Vice, F.L.; Gewolb, I.H. Variability of swallow-associated sounds in adults and infants. Dysphagia 2009, 24, 13–19.
- 27. Takahashi, K.; Groher, M.E.; Michi, K.-i. Methodology for detecting swallowing sounds. Dysphagia 1994, 9, 54–62.
- Taveira, K.V.M.; Santos, R.S.; Leão, B.L.C.d.; Stechman Neto, J.; Pernambuco, L.; Silva, L.K.d.; Canto, G.D.L.; Porporatti, A.L. Diagnostic validity of methods for assessment of swallowing sounds: A systematic review. Braz. J. Otorhinolaryngol. 2018, 84, 638–652.
- 29. Bi, Y.; Lv, M.S.; Song, C.; Xu, W.Y.; Guan, N.; Yi, W. AutoDietary: A Wearable Acoustic Sensor System for Food Intake Recognition in Daily Life. IEEE Sens. J. 2016, 16, 806–816.
- Fontana, J.M.; Melo, P.L.; Sazonov, E.S. Swallowing detection by sonic and subsonic frequencies: A comparison. In Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society, Boston, MA, USA, 30 August–3 September 2011; pp. 6890–6893.
- 31. Fukuike, C.; Kodama, N.; Manda, Y.; Hashimoto, Y.; Sugimoto, K.; Hirata, A.; Pan, Q.; Maeda, N.; Minagi, S. A novel automated detection system for swallowing sounds during eating and speech under everyday conditions. J. Oral Rehabil. 2015, 42, 340–347.
- 32. Kurihara, Y.; Kaburagi, T.; Kumagai, S.; Matsumoto, T. Development of Swallowing-Movement-Sensing Device and Swallowing-State-Estimation System. IEEE Sens. J. 2019, 19, 3532–3542.
- 33. Skowronski, M.D.; Crary, M.A.; Shrivastav, R. Acoustic discrimination of healthy swallows from upper airway movements. J. Acoust. Soc. Am. 2013, 134, EL127–EL132.
- Sejdic, E.; Steele, C.M.; Chau, T. Segmentation of Dual-Axis Swallowing Accelerometry Signals in Healthy Subjects with Analysis of Anthropometric Effects on Duration of Swallowing Activities. IEEE Trans. Biomed. Eng. 2009, 56, 1090–1097.
- Afkari, S. Measuring frequency of spontaneous swallowing. Australas. Phys. Eng. Sci. Med. 2007, 30, 313–317.
- Amft, O.; Troster, G. Recognition of dietary activity events using on-body sensors. Artif. Intell. Med. 2008, 42, 121–136.

- 37. Lee, J.; Steele, C.M.; Chau, T. Swallow segmentation with artificial neural networks and multisensor fusion. Med. Eng. Phys. 2009, 31, 1049–1055.
- 38. Makeyev, O.; Lopez-Meyer, P.; Schuckers, S.; Besio, W.; Sazonov, E. Automatic food intake detection based on swallowing sounds. Biomed. Signal Process. Control 2012, 7, 649–656.
- Sazonov, E.S.; Makeyev, O.; Schuckers, S.; Lopez-Meyer, P.; Melanson, E.L.; Neuman, M.R. Automatic Detection of Swallowing Events by Acoustical Means for Applications of Monitoring of Ingestive Behavior. IEEE Trans. Biomed. Eng. 2010, 57, 626–633.
- 40. Skowronski, M.D.; Harris, J.G. Exploiting independent filter bandwidth of human factor cepstral coefficients in automatic speech recognition. J. Acoust. Soc. Am. 2004, 116, 1774–1780.
- 41. Silva, J.; Chau, T. Coupled microphone-accelerometer sensor pair for dynamic noise reduction in MMG signal recording. Electron. Lett. 2003, 39, 1–2.
- 42. Khalifa, Y.; Coyle, J.L.; Sejdić, E. Non-invasive identification of swallows via deep learning in high resolution cervical auscultation recordings. Sci. Rep. 2020, 10, 8704.

Retrieved from https://encyclopedia.pub/entry/history/show/89517